


Introduction

- ▶ ASTER: Accélérateur pour les Sciences de la Terre, Environnement, Risques
- ▶ 5 MV Cockcroft-Walton by HVEE
- ▶ installation: October 2006
- ▶ acceptance tests: April 2007
- ▶ currently dedicated to Earth science studies
- ▶ 4 years of routine ¹⁰Be & ²⁶Al measurements
- ▶ ³⁶Cl [1], ⁴¹Ca & ¹²⁹I measurements



Main applications

- ▶ AMS developments
- ▶ cosmogenic nuclides systematics
- ▶ quantification of Earth's surface processes
- ▶ geochronology
- ▶ nuclear waste characterisation



Performance data (more information in [1,2])



	¹⁰ Be	²⁶ Al	³⁶ Cl	⁴¹ Ca	¹²⁹ I
extraction as	BeO ⁻	Al ⁻	Cl ⁻	CaF ₃ ⁻	I ⁻
currents [μA]	3-5	0.1-0.4	15-20	0.2-0.3	3-5
terminal voltage [MV]	4.5	2.7	5.0	5.0	5.0
stripped to (with absorber foil)	Be ²⁺ Be ⁴⁺	Al ³⁺	Cl ⁵⁺ Cl ¹⁰⁺	Ca ⁴⁺	I ⁵⁺
total transmission [%]	60	38	20	15	8
only absorber foil	36		14		
total with absorber foil	22		2		
detection efficiency [%]			56	68	
suppression factor (detector only)			³⁶ S 1 ³⁶ Cl 2.3·10 ⁶	⁴¹ K 1 ⁴¹ Ca 7·10 ³	
background [10 ⁻¹⁵]	0.15	0.9	0.28*	30	20
STDV [%] on 10 ⁻¹¹	0.3	1.3	1.0	1.6	1.1

*no AgBr backing! Ni cathodes with Ni pins



Machine layout








The life of ASTER



¹⁰Be, ²⁶Al: Statistics & Quality assurance




³⁶Cl & ¹²⁹I: Quality assurance



Résumé

- ▶ routine and fully automated measurements of ¹⁰Be & ²⁶Al sample
- ▶ after installation of new source head short-term sample to sample ³⁶Cl cross-contamination decrease >>> minimal measurement time losses for initial burn-in periods (5 min) for virgin targets and waiting periods between data acquisition of two sample runs (2 min)
- ▶ need for improved measurement strategy and data evaluation for accurate ³⁵Cl/³⁷Cl determination
- ▶ ¹²⁹I cross-contamination 0.5% (short-term) to 0.8% (long-term, up to 20 h measurement time)
- ▶ 3 secondary ¹²⁹I/¹²⁷I standards cross-calibrated vs. primary NIST 3231: SM-I-09 (1.007 ± 0.014)·10⁻⁹
SM-I-10 (1.064 ± 0.016)·10⁻¹⁰
SM-I-11 (1.083 ± 0.017)·10⁻¹¹
overall uncertainties (~1.5 %) mainly from primary NIST 3231 (at 10⁻⁸)
- ▶ quality assurance established >>> accuracy and reproducibility tested and improved >>> some remaining discrepancies with so far unknown reasons
- ▶ world-wide Earth science applications (see map)



References

- [1] R.C. Finkel et al., Improved ³⁶Cl Performance at the ASTER HVE 5 MV Accelerator Mass Spectrometer Facility, poster presentation @ AMS-12.
- [2] M. Arnold et al., NIMB 268 (2010) 1954.
- [3] S. Merchel et al., Ultra-trace analysis of ³⁶Cl by accelerator mass spectrometry: An interlaboratory study, submitted to Anal. Bioanal. Chem.

Acknowledgments

- ▶ I. Mrak for Slovenian collaboration (Veliki Vrh)
- ▶ all ASTER users and HVEE for excellent collaborations
- ▶ W. Bremser & colleagues from other AMS facilities for round-robin exercises
- ▶ for cash: DAAD

